

EC DECLARATION OF CONFORMTLY

I.A.C.E.R. S.r.l

Via Enzo Ferrari 2 – 30037 Scorzè (Ve), Italia

herewith declares under its own responsibility, that the product

I-TECH REACARE

UMDNS Code: 11244 Batch no.: Series no.:

complies with the essential safety requirements of the European Medical Device Directive 93/42/EEC (transposed in Italy by the D.Lgs. 46/97), as modified by the Directive 2007/47/EC (D.Lgs.37/2010) and further modifications/integrations.

The product has been assigned to class IIb, according to Annex IX, rule 9 of the Directive 93/42/EEC (and further modifications/integrations) and bears the mark



Compliance of the concerned product with the Directive 93/42/EEC has been assessed and certified by the **Notified Body**

> 0068 - MTIC InterCert S.r.l. Via G. Leopardi 14, Milano (MI) 20123

> Certificate no.: 0068/QCO-DM/168-2020

following the certification procedure according to Annex II (excluding point 4) of the Directive 93/42/EEC.

The devices comply with the following applicable standards:

EN 60601-1:2006 + A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010 + A1:2015, EN ISO 14971: 2019, ISO/TR 24971:2020, ISO 10993-1: 2018, ISO 10993-5: 2009, ISO 10993-10: 2010, EN 62304: 2006, EN 62366-1: 2015.

It is also claimed that:

- The devices do not incorporate, as an integral part, any substance or a human blood derivate of point 10 of Annex 1 (Directive 2007/47/CE);
- No fabrics of animal origin have been used in the production as per Regulation (EU) no. 722/2012 of the Commission.

Scorzè, 31/01/2023

Place, date

MASSIMO MARCON

Legal Representative

MD117-07 Date.Rev.31/01/22





