

This Declaration of Conformity is issued under the sole responsibility of the manufacturer. The device covered by the present Declaration is in conformity with all regulations or directives below, including compliance with related Essential Requirements and General Safety and Performance Requirements.

### 1. Object of the declaration

<b>Product Name</b>	Infrared Thermometer
<b>Model Number</b>	AOJ-20A
<b>Product Type</b>	Medical Thermometer
<b>Intended Purpose</b>	The digital thermometer provides a quick and highly accurate reading of an individual's body temperature.
<b>Product Descriptions</b>	The digital thermometer is intended to measure the human body's temperature in regular mode orally, rectally or under the arm, and the device is reusable for clinical or home use on people of all ages.
<b>Basic UDI-DI</b>	697204011AOJ25X17Z
<b>Control Indicator</b>	Lot number

The object of the Declaration described above is in conformity with the following regulations:

<b>EU Directive</b>	<b>Medical Device Directive (93/42/EEC as amended by 2007/47/EC)</b>
<b>Device Risk Classification</b>	Class IIa based on Annex V
<b>Conformity Assessment Path</b>	Annex IX Conformity assessment based on a quality management system and on assessment of technical documentation
<b>Notified Body Name, Address, and ID</b>	<b>NB Name:</b> TÜV SÜD Product Service GmbH <b>Address:</b> Ridlerstraße 65, 80339 MÜNCHEN, Germany <b>NB Code:</b> 0123
<b>Certificate(s) issued</b>	<b>G2 1037030001 Rev.00</b>
<b>Standards</b>	The products listed on this Declaration of Conformity have been assessed and/or tested in a typical configuration as described in the Manufacturer's accompanying documentation in accordance with the product standards listed below.  ISO 14155:2020, ISO 14971:2019, IEC 60601-1:2005,AMD:2012, IEC 60601-1-11:2015 for use in conjunction with IEC 60601-1:2005,AMD:2012,ISO 80601-2-56:2017,AMD1:2018 for use in conjunction with IEC 60601-1:2005,COR1:2006,COR2:2007,AMD1:2012,IEC 60601-1-2:2014,EN 60601-1-2:2015,IEC 60601-1-11:2015 clause 12, EN 60601-1-11:2015 clause 12, ISO 80601-2-56:2017+A1:2018 clause 201.17&202, EN ISO 80601-2-56:2017+A1:2020 clause 201.17&202, ISO 10993-5:2009,ISO 10993-10:2010,

<b>EU Directive</b>	<b>RoHS Directive EU) 2015/863</b>
---------------------	------------------------------------

<b>EU Directive</b>	<b>Directive 2011/65/EU of the European Parliament and of the Council of 08 June 2011 on the restrictions of the use of certain hazardous substances in electrical and electronic equipment, amended up to and inclusive of Directive (EU) 2017/2102 (RoHS)</b>
<b>Device Classification</b>	Category 8, medical device, according to Annex I
<b>Standards</b>	The products listed on this Declaration of Conformity have been assessed and/or tested in a typical configuration as described in the Manufacturer's accompanying documentation in accordance with the product standards listed below. IEC 62474:2012, IEC 62321:2013, EN 62321:2009, EN 50581:2012, IEC/TR62476

## 2. Additional information

<b>Manufacturer</b>	<b>Name:</b> Shenzhen AOJ Medical Technology Co., Ltd. <b>Address:</b> Room 301&4F, Block A, Building A, Jingfa Intelligent Manufacturing Park, Xiaweiyuan, Gushu Community, Xixiang Street, Bao'an District, 518126, Shenzhen, China <b>SRN:</b> CN-MF-000018386
<b>EU Authorized Representative</b>	<b>Name:</b> Share Info GmbH <b>Address:</b> Heerdter Lohweg 83, 40549 Düsseldorf. <b>SRN:</b> DE-AR-000005132
<b>Quality Certificates Issued</b>	The Manufacturer is certified by TUV to the following: EN ISO 13485:2016 , as evidenced by certificate number Q5 103703 0004 Rev.00

Signature (signed for and on behalf of Shenzhen AOJ Medical Technology Co., Ltd.):

Date of Issue:

Printed Name: Baiyu Xie  
Title: Regulatory Assistant

Place of Issue: Shenzhen

Jack Wang  
28/10 2022